FUTEK MODEL LLB300

SUBMINIATURE LOAD BUTTON LOAD CELL

(Previously L1630-Q10291)

Drawing Number: FI1119-A	
INCH [mm]	R.O.= Rated Output

WIRING CODE (WC1)			
+Excitation	-Excitation	+Signal	-Signal
RED	BLACK	GREEN	WHITE
SHIELD			
FLOATING			







STK#	CAPACITY lb	N
FSH01045	10	44.5
FSH01046	25	111
FSH01047	50	222
FSH01048	100	445
FSH01050	250	1112
FSH01052	500	2224
FSH01053	1000	4448

SPECIFICATIONS:

RATED OUTPUT	*2 mV/V nom.*
SAFE OVERLOAD	*150% of R.O.*
ZERO BALANCE	*±3% R.O.*
EXCITATION (VDC OR VAC)	*18 MAX*
BRIDGE RESISTANCE	*700 Ω nom.*
NONLINEARITY	*±0.5% of R.O.* } IMPROVED ACCURACY
HYSTERESIS	*±0.5% of R.O.* } AVAILABLE CONTACT
NONREPEATABILITY	*±0.1% of R.O.* } FACTORY
TEMP. SHIFT ZERO	*±0.01% of R.O./°F [0.018% of R.O./°C]*
TEMP. SHIFT SPAN	*±0.02% of LOAD/°F [0.036% of LOAD/°C]*
COMPENSATED TEMP.	*60 to 250°F [15 to 121°C]*
OPERATING TEMP.	*-60 to 250°F [-50 to 121°C]*
WEIGHT	*1.5 oz [43 g]*
MATERIAL	*17-4PH S.S.*
DEFLECTION	*0.002 [0.05] nom.*

CABLE: #29 AWG, 4 Conductor, Spiral Shielded Teflon Cable 10 ft [3m] Long

ACCESSORIES AND RELATED INSTRUMENTS AVAILABLE

CALIBRATION (STD)	*5 pt. COMPRESSION; 100K Ω SHUNT CAL. VALUE*
CALIBRATION TEST EXCITATION	*10 VDC*

FUTEK
ADVANCED SENSOR TECHNOLOGY, INC.

This drawing is submitted solely for the information and exclusive use of the original addressee. It is not to be divulged in whole or in part, by any firm or individual without written permission from FUTEK

10 THOMAS
IRVINE, CA 92618 USA
1-800-23-FUTEK (38835)

INTERNET:
http://www.futek.com